

Surface Electrode

NM-315S

General

This electrode is designed for detecting signals in the evoked EMG examination.

Safety Information

⚠ WARNING A warning alerts the user to possible injury or death associated with the use or misuse of the instrument.

⚠ CAUTION A caution alerts the user to possible injury or problems with the instrument associated with its use or misuse such as instrument malfunction, instrument failure, damage to the instrument, or damage to other property.

Pay attention to all safety information in the Operator's Manual or Installation Guide.

⚠ WARNING

Connect the lead tip to the specified input jack only. Otherwise, the patient could receive an electrical shock.

⚠ WARNING

Do not use this electrode during MRI examination because it may cause skin burn on the electrode placement area.

⚠ CAUTION

Do not attach the electrode on a wound or inflamed skin.

⚠ CAUTION

Do not use this electrode for electric simulation.

⚠ CAUTION

Do not store this electrode in an atmosphere with sulfur. It will deteriorate the electrode.

⚠ CAUTION

Do not clean the electrode with hot water or chloric bleaching agent.

⚠ CAUTION

- Do not sterilize the electrode.
- Do not disinfect the electrode with disinfecting alcohol, cresol or glutaraldehyde.

This Safety and Performance Information is an extract from the general and safety information sections of the most recent edition of Operator's Manual or Installation Guide. Therefore, the contents of your Operator's Manual or Installation Guide may differ from those of this Safety and Performance Information. For detailed operating procedures, follow the instructions of your Operator's Manual or Installation Guide.

 Manufacturer
NIHON KOHDEN CORPORATION
 1-31-4 Nishiochiai, Shinjuku-ku,
 Tokyo 161-8560, Japan
 Phone +81 3-5996-8041
<https://www.nihonkohden.com/>

EC REP European Representative
NIHON KOHDEN EUROPE GmbH
 Raiffeisenstrasse 10, 61191 Rosbach, Germany
 Phone +49 6003-827-0 Fax +49 6003-827-599

1st Edition: 14 May 2019
 3rd Edition: 12 Dec 2022

