


**MK-RE3 HEMATOLOGY CONTROL**  
for reticulocyte **3.0mL**

**ASSAY SHEET**

<b>LOT</b>	MKRE247-3
	2024-09-05
<b>OPEN</b>	14 DAYS

		<b>MEK-9200</b>	<b>RANGE</b>
<b>RET%</b>	<b>%</b>	<b>7.78</b>	<b>±4.22</b>
<b>RET</b>	<b>10<sup>6</sup>/μL</b>	<b>0.2807</b>	<b>±0.1522</b>

**2D Barcode Entry**



• If RBC value is under 3.00 [ $\times 10^6/\mu\text{L}$ ], please refrain from using.  
It may have deteriorated or hemolyzed, which affects the measurement accuracy.



<https://www.nihonkohden.com/spi/index.html>

<b>Manual Registration</b>						
Lot	247					
Level	3					
Expiration	9	M	5	D	2024	Y
Control	MKRE					



**NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN