

MEK-5DH HEMATOLOGY CONTROL
for WBC 5 part differential 3.0mL

ASSAY SHEET

LOT	MEK5D248H
	2024-11-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry



		RANGE	MEK -7222	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10 ³ /μL	±2.2	24.8	24.7	24.3	25.56
NE%	%	±10.0	75.1	75.1	75.1	76.49
LY%	%	±10.0	11.5	11.5	11.5	12.81
MO%	%	±5.0	9.7	9.7	9.7	7.03
EO%	%	±5.0	2.8	2.8	2.8	2.76
BA%	%	±4.0	0.9	0.9	0.9	0.91
NE	10 ³ /μL	±2.8	18.6	18.6	18.2	19.55
LY	10 ³ /μL	±2.8	2.9	2.8	2.8	3.27
MO	10 ³ /μL	±1.0	2.4	2.4	2.4	1.80
EO	10 ³ /μL	±1.0	0.7	0.7	0.7	0.71
BA	10 ³ /μL	±0.8	0.2	0.2	0.2	0.23
RBC	10 ⁶ /μL	±0.25	5.81	5.77	5.76	5.88
HGB	g/dL	±0.7	18.3	18.0	18.1	18.12
HCT	%	±4.0	57.7	56.1	54.5	65.2
MCV	fL	±8.0	99.3	97.2	94.6	110.8
MCH	pg	±2.8	31.5	31.2	31.4	30.8
MCHC	g/dL	±3.0	31.7	32.1	33.2	27.8
RDW-CV *	%	±4.0	13.5	14.2	13.6	15.4
PLT	10 ³ /μL	±70	568	559	566	576.5
MPV	fL	±4.0	8.1	8.3	8.2	10.3

Manual Registration	
Lot	248
Level	High
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-7222,8222.



<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



0694-901692P